



# EU Quality Management System Certificate

Regulation (EU) 2017/745 on Medical Devices, Annex IX Chapter I

**Certificate No. G15 122977 0002 Rev. 00**

**Manufacturer:**

**Preci Health SA**

Rue du Prébarreau 17  
2000 Neuchâtel  
SWITZERLAND

SRN Manufacturer - CH-MF-000045035

**Authorized  
Representative:**

MedNet EC-REP GmbH  
Borkstraße 10, 48163 Münster, GERMANY

The quality management system has been evaluated in accordance with Regulation (EU) 2017/745, Annex IX Chapter I with a positive result.

Details on devices covered by the quality management system are described on the following page(s). The report referenced below summarises the results of the assessment and includes reference to relevant CS, harmonised standards and test reports.

The certified quality management system is subject to periodical surveillance.

If class I devices in sterile conditions, with measuring function, or reusable surgical instruments are covered by this certificate, the audit was limited to the respective aspects relating to

- establishing, securing, and maintaining sterile conditions,
- conformity of the devices with the metrological requirements,
- reuse of the device, in particular cleaning, disinfection, sterilization, maintenance and functional testing and the related instructions for use.

If class IIa or class IIb devices are covered by this certificate, the quality management system assessment was accompanied by the assessment of technical documentation for devices selected on a representative basis. The periodical surveillance includes further assessment of the technical documentation on the basis of representative samples.

If class III (excluding custom-made implantable devices) or class IIb implantable devices are covered by this certificate, an EU Technical Documentation Assessment Certificate in accordance with Annex IX Chapter II is required before placing them on the market.

All applicable requirements of the Testing, Certification, Validation and Verification Regulations TÜV SÜD Group have to be complied with.

For details and certificate validity see: [www.tuvsud.com/ps-cert?q=cert:G15 122977 0002 Rev. 00](http://www.tuvsud.com/ps-cert?q=cert:G15 122977 0002 Rev. 00)

**Report No.:** 713369476

**Valid from:** 2026-01-26

**Valid until:** 2031-01-25

Christoph Dicks  
Head of Certification/Notified Body

**Issue date:** 2026-01-26



Benannt durch/Designated by  
Zentralstelle der Länder  
für Gesundheitsschutz  
bei Arzneimitteln und  
Medizinprodukten  
www.zlg.de  
BS-MDR-099



Product Service

# EU Quality Management System Certificate

Regulation (EU) 2017/745 on Medical Devices, Annex IX Chapter I

**Certificate No. G15 122977 0002 Rev. 00**

**Classification:**

Class IIa

**Device Group:**

MDN 1202 - Non-active non-implantable devices for administration, channelling and removal of substances, including devices for dialysis

**The validity of this certificate  
depends on conditions and/or  
is limited to the following:**

./.

## Revision History:

Rev.	Dated	Report	Description
00	2026-01-26	713369476	Initial issuance